Pressure Indicator Gauge Series 1500 81/2" Dial

WIKA Datasheet 1500 8.5I

Applications

 Suitable for test and laboratory and production applications

Special features

Capsule-type pressure elements up to and including 150 psig, Bourdon tubes for ranges of 200 psig and above

Tempered-glass window, solid-front dial and a blow-out plug on the back

Available in 26 standard ranges (including vacuum and compound ranges)

■ Also available installed in a compact, suitcase-type carrying case with a mounted panel for the gauge. Cover is easily removed and connections can be made without removing the gauge from the case.



Pressure Indicator Gauge Series 1500 - 8.5" Dial

Standard features

Accuracy 0.066% of full scale

Range -30" Hg -1,000 psig

Repeatability

0.03% of full scale

Sensitivity 0.01% of full scale

Hysteresis

0.1% of full scale for ranges of 150 psi and below; 0.2% of full scale for 200 psi and above

Maximum temperature effect

0.1% of full scale per 10°C/18°F change from reference temperature 23°C/73.4°F

Scale length

Approx. 45" through 2 pointer revolutions

Dial diameter 81/2"

Pressure element volume

Ranges <150 psi 6.9 cc with pointer at zero 8.6 cc At full scale ranges 200 psi and above 10 cc. For the range 0-125" water only 6.3 cc with pointer at zero; 8.0 cc, at full scale

Case pressure

Cases are open to atmosphere.

Case connections

150 psi and below, $\frac{1}{8}$ " female NPT, 200 psi and above, $\frac{1}{4}$ " female NPT built-in stainless steel filter at the bottom of case

Case construction

Aluminum with tempered-glass window, flush mounted

Zero adjustment

Accessible through both the front and the back of the case

Options

- Calibration in most metric units available at no extra cost. Other calibration units and dual scale graduations are available at extra cost.
- Gauge 1500 8.5" is available installed in a compact, suitcase-type carrying case with a mounted panel for the gauge. The cover is easily removed and connections can be made without removing the gauge from the case

Note: Gases must be non-corrosive, no liquid service.

Weight and shipping weight 12 lbs.

2 103.

Ordering information

Please specify ordering number, range, and mounting angle, (extra cost if mounting angle other than vertical)

Page 1 of 2

Part of your business

WIKA Datasheet 1500 8.5I 02/2014

Series 1500 8¹/₂" Pressure Indicator Gauge

Wide Spaced Graduations Give Excellent Readability

Because the pointer covers full scale in two revolutions, Series 1500 scales are approximately 45" long. This is more than twice as long as single-revolution gauges with the same dial diameter ($8\frac{1}{2}$ ") and longer than gauges with twice the dial diameter. The expanded scale allows a minimum of 0.045" of white space between graduations.

Accurate Readout at a Glance

Compactness makes readout convenient as well as accurate. The operator takes readings from one position, which is not possible with large-diameter gauges. This cuts fatigue and errors. A revolution indicator is included and a mirror ring eliminates parallax errors. Graduations are easily seen; bold numbers are horizontally placed.

Calibration is Traceable to National Institute of Standards and Technology (NIST)

A computer-assisted plotter marks calibration points and the graduations between them on each dial. This produces a scale which precisely matches the characteristics of each mechanism and pressure element. Instruments supplied are certified traceable to NIST.

Compact Size Saves Space in Panels

With the two-revolution arrangement, the long scale (approx. 45") requires a dial only 8½" in diameter, making more compact test units possible, such as consoles for ground support equipment.

Performs Better than the Rated Accuracy

Custom dial calibration, and individual assembly and adjustment of each mechanism add up to an accuracy of 0.066% full scale. This figure is the minimum performance which can be expected. After rigorous testing any WIKA gauge which fails to better the rated accuracy is rejected.

Solid-Front Construction

The case has a heavy aluminum solid-front and a blowout plug in back of the case. The pressure inlet is only $1/_{16}$ " in diameter; the blow-out plug is $1\frac{1}{2}$ ". This design gives excellent protection against excessive pressure as the plug relieves at the slightest build-up. The case is heavy cast aluminum with a tempered glass window. Overpressuring the instrument up to 10% above its full scale range will not damage the mechanism nor affect accuracy. Series 1500 81/2" Pressure Indicator Gauge

Standard Ranges and Ordering Numbers

| Range and Calibration | Ordering Number | Graduation |
|-----------------------|--------------------|------------|
| 0-4.5 psi | 62A-2A-0005 | 0.005 psi |
| 0-10 psi | 62A-2A-0010 | 0.01 psi |
| 0-15.5 psi | 62A-2A-0015 | 0.02 psi |
| 0-20 psi | 62A-2A-0020 | 0.02 psi |
| 0-30 psi | 62A-2A-0030 | 0.05 psi |
| 0-45 psi | 62A-2A-0045 | 0.05 psi |
| 0-60 psi | 62A-2A-0060 | 0.1 psi |
| 0-100 psi | 62A-2A-0100 | 0.1 psi |
| 0-150 psi | 62A-2A-0150 | 0.2 psi |
| 0-200 psi | 62A-2A-0200 | 0.2 psi |
| 0-300 psi | 62A-2A-0300 | 0.5 psi |
| 0-500 psi | 62A-2A-0500 | 0.5 psi |
| 0-1,000 psi | 62A-2A-1000 | 1.0 psi |

| Range and Calibration | Ordering Number | Graduation |
|-----------------------|--------------------|------------|
| 0-125" water | 62A-2C-0125 | 0.2" |
| 0-280" water | 62A-2C-0280 | 0.5" |
| 0-800 mm Hg | 62A-2D-0800 | 1.0 mm |
| 0-31.5" Hg | 62A-2B-0031 | 0.05" |
| 0-60" Hg | 62A-2B-0060 | 0.1" |
| 0-120" Hg | 62A-2B-0120 | 0.2" |
| 0-200" Hg | 62A-2B-0200 | 0.2" |
| 0-300" Hg | 62A-2B-0300 | 0.5" |

Vacuum and Compound Ranges

| Range | Ordering Number | Graduation |
|---------------------|--------------------|------------|
| -30" Hg to 0 | 62A-3B-0030* | 0.05" |
| 15 to 0 to 15" Hg | 62A-5B-0030* | 0.05" |
| -30 to 0 to 30" Hg | 62A-5B-0060* | 0.1" |
| 30 to 0 to 90" Hg | 62A-5B-0120* | 0.2" |
| -30 to 0 to 270" Hg | 62A-5B-0300* | 0.5" |

*Can be calibrated as shown, or in any other standard pressure unit at no extra cost. Specify if other than in Hg.

WIKA Datasheet 1500 8.5I 02/2014



